

MANOMÈTRES TRANSPORTABLES À COLONNE DE LIQUIDE VERTICALE

Série KM

Pression / Dépression



KEY POINTS

The KM vertical liquid column portable manometer, developed and manufactured by KIMO, measures low pressures in gas networks.

- Easy to carry.
- "U"- shaped column for pressure and depression measurement.
- Direct read-off by moving the graduated slide strip.
- Safety valves actuation for momentaneous overshooting of the scale.
- Fitted with valve connectors and mounting hook.
- Supplied with connection sleeves, a bottle of VOLT 1S liquid and carrying case.

MEASURING RANGE

Reference	Measuring range	Resolution
KM 45	0 – 45 mbar	0.2 mbar
KM 60	0 – 60 mbar	0.2 mbar

DIMENSIONS


Reference	Dimensions (height. x width x thickness)	Weight (with accessories)
KM 45	306 X 50 X 20 mm	550 g
KM 60	380 X 50 X 20 mm	700 g

TECHNICAL FEATURES

Recommended range of use	From +5 to +30°C
Possible range of use	From -30 to +60°C
Pression statique maximum	8 bars
Manometer body	15 mm thick transparent Altuglas.
Liquid column	Ø 4 mm bored into the solid block.
Graduated slide strip	Altuglas transparent
Zero adjustment	By moving the graduated slide strip Fixed in place via milled, nickel-plated brass screw.
Manometric liquid	VOLT 1S, density 1.86 at 20°C.
Connection	On Ø 6.2 nickel-plated brass valve connectors, with 1m long neoprene tube fitted with dedicated end-pieces for gas equipment.

MOUNTING

1. **Dismount** one of the 2 connectors using a no.12 spanner and slacken the milled head of the other connector by one turn.
2. Check beforehand that the **slide strip** is at its **lowest level**.
3. **Pour** the liquid in the column using the spout.
4. **Do not overfill. Never go beyond the NL line at the middle of the slide strip.**
5. **Remount** the connector and screw the milled head of the other connector back down.



OPERATION

1. **Hang up** the manometer vertically by the mounting hook or hold manually.
2. **Open** to the air by slackening the milled heads of the 2 valves (one turn is sufficient).
3. **Push one** of the ends of the connecting tube firmly onto the right-hand valve. Push the other end of the tube onto the pressure point of the pipeline or the instrument which has to be checked.
4. Le liquide, sous la pression du gaz, baisse dans la colonne de droite et monte dans celle de gauche. **Si l'arrivée du gaz s'opère trop brutalement et fait jouer le clapet de sécurité**, recommencer l'opération en pinçant le tube de liaison plus ou moins fortement pour admettre le gaz plus lentement (Si à nouveau le clapet de sécurité fonctionne, c'est que la pression à contrôler dépasse largement l'étendue de mesure du manomètre).
5. When the liquid has settled, **slide the graduated strip** so as to bring the zero mark opposite the right-hand tube's liquid level (lowest level).
6. The graduation corresponding to the **height of the liquid in the left-hand tube** indicates the exact gas pressure.
7. Close off the **2 valves securely** after operation

IMPORTANT :

- Only **VOLT1S liquid** will ensure precise measurement (slide scale graduation corresponding to the density of this liquid).
- Maximum static pressure : **8 bars**

www.kimo.fr

Distributed by :



EXPORT DEPARTMENT

Tel : + 33. 1. 60. 06. 69. 25 - Fax : + 33. 1. 60. 06. 69. 29

e-mail : export@kimo.fr